



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/749,270	12/31/2003	Wayne A. Becker	1001.1418102	8516
28075	7590	08/09/2005		
CROMPTON, SEAGER & TUFTE, LLC 1221 NICOLLET AVENUE SUITE 800 MINNEAPOLIS, MN 55403-2420				
			EXAMINER	
			MENDOZA, MICHAEL G	
			ART UNIT	PAPER NUMBER
			3731	

DATE MAILED: 08/09/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/749,270

Applicant(s)

BECKER ET AL.

Examiner

Michael G. Mendoza

Art Unit

3731

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 11 May 2005.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 20-42 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 20-42 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____.
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____.

Response to Arguments

1. Applicant's arguments filed 11 May 2005 have been fully considered but they are not persuasive. The Applicant argues that Becker fails to teach a delivery sheath. The Applicant states argues that the inflation balloon and not the catheter shaft increases in outer dimension. The device of Becker is taken as a whole therefore the balloon of Becker increases the overall outer dimension of the device. The Applicant argues that Becker does not teach a delivery sheath. The device of Becker as a whole is a delivery sheath. Becker teaches a lumen that is fully capable of transporting an intravascular filter. Furthermore a recitation of the intended use of the claimed invention must result in a structural difference between the claimed invention and the prior art in order to patentably distinguish the claimed invention from the prior art. If the prior art structure is capable of performing the intended use, then it meets the claim. In a claim drawn to a process of making, the intended use must result in a manipulative difference as compared to the prior art. See *In re Casey*, 370 F.2d 576, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 312 F.2d 937, 939, 136 USPQ 458, 459 (CCPA 1963).

Terminal Disclaimer

2. The terminal disclaimer filed on 11 May 2005 disclaiming the terminal portion of any patent granted on this application which would extend beyond the expiration date of 6689151 has been reviewed and is accepted. The terminal disclaimer has been recorded.

Claim Rejections - 35 USC § 102

3. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

4. Claims 20-42 are rejected under 35 U.S.C. 102(b) as being anticipated by Becker et al. 6165152.

5. Becker et al. teaches a proximal region having a proximal end and a distal end; an intermediate region extending distally from the proximal region, the intermediate region increasing in outer dimension from a proximal end of the intermediate region to a distal end thereof; a distal portion extending distally from the intermediate region, the distal portion including a proximal end having an outer dimension that is greater than the outer dimension of the distal end of the intermediate region; wherein the distal portion distally tapers in wall thickness along at least a portion of its length; wherein the proximal region of the delivery sheath has a constant outer dimension; wherein the distally tapering wall thickness is formed by distally decreasing the outside dimension of the delivery sheath (col. 5, lines 27-36); wherein the distal portion includes a distal region, a reduced diameter region, and a distal tip region; and wherein the distal tip region is a closed distal tip region (col. 5, lines 16-17).

6. As to claims 25, 31, 36, and 42 the function recitation that the delivery sheath is configured to accommodate an intravascular emboli capturing device therein has not been given patentable weight because it is narrative in form. In order to be given

Art Unit: 3731

patentable weight, a function recitation must be expressed as a "means" for performing the specified function, as set forth in 35 USC § 112, 6TH paragraph, and must be supported by recitation in the claim of sufficient structure to warrant the presence of the function language. *In re Fuller*, 1929 C.D. 172; 388 O.G. 279.




FIG. 1




FIG. 2

Conclusion

7. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Contacts

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael G. Mendoza whose telephone number is (571) 272-4698. The examiner can normally be reached on Mon.-Fri. 8:00 a.m. - 5:00 p.m..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Anh Tuan Nguyen can be reached on (571) 272-44963. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3731

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



MM



GLENN K. DAWSON
PRIMARY EXAMINER